




Test Certificate

Part 1 Liter Vessel
Part Number 02503
Serial Number 91506 - 91544



		USP / Ph.Eur. Tolerances	
Dimensions	Vessel	Inside Ø a	100 mm 98 – 106 mm
		Height b	200 mm 160 – 210 mm
		With hemispherical bottom	
Material	Borosilicate glass		

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopoeia requirements (at date of approval, see below).

Aesch, 2022-10-24

Bruno Balli
SOTAX Quality Management
SOTAX AG | Nording 1 | 4147 Aesch | Switzerland