



# Test Certificate

Part **Calibrated Measuring  
Cylinder 250 mL**  
Part Number **08425**



		USP / Ph.Eur. Tolerances	
<b>Dimensions</b>	Cylinder	Total height	not more than 335 mm
		Graduated	not less than 200 mm
		Readable	± 2 mL
		Weight	220 g ± 44 g
<b>Material</b>	Borosilicate Glass (Duran)		

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopoeia requirements (at date of approval, see below).

Aesch, 2019-11-06

Bruno Balli  
SOTAX Quality Management  
SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland