




Test Certificate

Part 1 Liter Vessel
Part Number 14143-01
Serial Number 92573 - 92724



| | | USP / Ph.Eur. Tolerances | | |
|-------------------|--------------------|---------------------------|-------------------|--------------|
| Dimensions | Vessel | Inside Ø a | 102.9 to 105.1 mm | 98 – 106 mm |
| | | Height b | 200 mm | 160 – 210 mm |
| | | With hemispherical bottom | | |
| Material | Borosilicate glass | | | |

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at date of approval, see below).

Aesch, 2022-12-02

Bruno Balli
SOTAX Quality Management
SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland