


Test Certificate

Part: **1 Liter Vessel**
 Number: 14322-01
 Serial Number: 72316 - 72365

For SOTAX AT Dissolution Testing Systems



Dimensions:	Inside Ø	a	98.5 to 101.5 mm	USP/Ph.Eur. Tolerances
	Height	b	200 mm	
	With hemispherical bottom			
Material:	Borosilicate glass			

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management
 Bruno Balli
 2017-10-18


 Signature