

Test Certificate

Part: **Paddle with Suction Opening**
 Number: 02985-2
 Serial Number: 19001 - 19250

For SOTAX AT Dissolution Testing Systems



				USP/Ph.Eur. Tolerances
Dimensions:	Paddle	Height	19.0 mm	± 0.5 mm
		Upper length	74.5 mm	± 0.5 mm
		Lower length	42.0 mm	± 1.0 mm
		Size	4.0 mm	± 1.0 mm
		Radius	41.5 mm	± 1.0 mm
	Shaft	Diameter	9.4 to 10.1 mm	- -

Material: **Stainless Steel (electrolytic polished)**

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management
 Bruno Balli
 2013-05-15


 Signature