




Part **Beaker**
 Part Number **03061**



Test Certificate

**USP / Ph.Eur.
Tolerances**

| | | | | |
|-------------------|--------------------|----------------------|-------------------|--------------|
| Dimensions | Beaker | Inside Ø a | 104.5 to 107.5 mm | 97 – 115 mm |
| | | Height b | 145 mm | 138 – 160 mm |
| | | With straight bottom | | |
| Material | Borosilicate glass | | | |

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at date of approval, see below).

Aesch, 2019-11-05

Bruno Balli
 SOTAX Quality Management
 SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland