


## Test Certificate

Part 1 Liter Vessel  
 Part Number 14143-01  
 Serial Number 104598 - 104943



		USP / Ph.Eur. Tolerances	
Dimensions	Vessel	Inside Ø a	102.9 to 105.1 mm
		Height b	200 mm
With hemispherical bottom			
Material	Borosilicate glass		

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopoeia requirements (at date of approval, see below).

Aesch, 2025-11-26

  
 05.04.2028 09:34  
 Buner Cornelius  
 SOTAX Quality Management  
 SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland