

± 2.0 mm ± 1.5 mm Distance to sieve 8.2 mm 3 Holes Ø 38.3 mm ± 2.3 mm PMMA, POM, PVC, PA Material

USP / Ph.Eur conformity (acc. USP701 / Ph Eur 2.9.1) :

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed.

The design of the basket-rack assembly may be varied somewhat, provided the specifications for the glass tubes and the screen mesh size are maintained. (separate Certificates for Glass tube #8825-1 and Sieve plate # 23443-01)

Aesch, 2023-09-26

Michael Schmidlin SOTAX Quality Management SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland

1/1

CT00000EN /