


Test Certificate

Part: **1 Liter Vessel**
 Number: **02503**
 Serial Number: **56355 - 56605**

For SOTAX AT Dissolution Testing Systems



				USP/Ph.Eur. Tolerances
Dimensions:	Inside Ø	a	100 mm	98 – 106 mm
	Height	b	200 mm	160 – 210 mm
	With hemispherical bottom			
Material:	Borosilicate glass			

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management
 Bruno Balli
 2014-03-05



Signature