

Test Certificate

Part: **Basket Holder with Suction Opening**
Number: 05301
Serial Number: 32001 - 32250

For SOTAX AT Dissolution Testing Systems



Dimensions:	Cap Ø	25.0 mm	USP/Ph.Eur. Tolerances
	Cap size	5.1 mm	± 3.0 mm
	Clear opening	20.2 mm	± 0.5 mm
	Vent hole Ø	2.0 mm	± 1.0 mm
Material:	Stainless Steel		± 0.5 mm

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management
Bruno Balli
2014-07-17


Signature