

# Test Certificate

Part: **Basket Holder with Stirring Shaft**  
 Number: 07558  
 Serial Number: 32251 - 32500

For SOTAX AT Dissolution Testing Systems



			USP/Ph.Eur. Tolerances
Dimensions:	Shaft Ø	9.4 to 10.1 mm	- -
	Cap Ø	25.0 mm	± 3.0 mm
	Cap size	5.1 mm	± 0.5 mm
	Clear opening	20.2 mm	± 1.0 mm
	Vent hole Ø	2.0 mm	± 0.5 mm
Material:	Stainless Steel		

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management  
 Bruno Balli  
 2014-10-09

  
 Signature