


Test Certificate

Part: **Calibrated Measuring
Cylinder 100 mL**

Number: 08424

For SOTAX TD Tap Density Tester



Dimensions:

Total height
Graduated
Readable
Weight

130 g

USP/Ph.Eur.
Tolerances

not more than 335 mm
not less than 200 mm
 ± 1 mL
 ± 16 g

Material:

Borosilicate Glass (Duran)

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management
Bruno Balli
2015-11-12


Signature