


Test Certificate

Part: **1 Liter Vessel**
 Number: **14322-01**
 Serial Number: **75700 - 75741**

For SOTAX AT Dissolution Testing Systems



Dimensions:

Inside Ø a 98.5 to 101.1 mm
 Height b 200 mm

USP/Ph.Eur.
 Tolerances
 98 – 106 mm
 160 – 210 mm

With hemispherical bottom

Material:

Borosilicate glass

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at time of delivery).

SOTAX Quality Management
 Bruno Balli
 2018-08-30


 Signature